

Kansas Department on Aging

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION		(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER: N089001	(X2) MULTIPLE CONSTRUCTION A. BUILDING: _____ B. WING: _____	(X3) DATE SURVEY COMPLETED 02/25/2015
NAME OF PROVIDER OR SUPPLIER BREWSTER HEALTH CENTER		STREET ADDRESS, CITY, STATE, ZIP CODE 1001 SW 29TH ST TOPEKA, KS 66611		
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (EACH DEFICIENCY MUST BE PRECEDED BY FULL REGULATORY OR LSC IDENTIFYING INFORMATION)	ID PREFIX TAG	PROVIDER'S PLAN OF CORRECTION (EACH CORRECTIVE ACTION SHOULD BE CROSS-REFERENCED TO THE APPROPRIATE DEFICIENCY)	(X5) COMPLETE DATE
S 000	INITIAL COMMENTS The following citations represent the findings of a Health Resurvey.	S 000		
S 970 SS=F	26-40-302 (g)(i)(ii)(iii) P E - Nursing facility support systems (G) If a nursing facility uses a wireless system to meet the requirements of paragraphs (i)(1)(A) through (E), all of the following additional requirements shall be met: (i) The nursing facility shall be equipped with a system that records activated calls. (ii) A signal unanswered for a designated period of time, but not more than every three minutes, shall repeat and also be sent to another workstation or to staff that were not designated to receive the original call. (iii) Each wireless system shall utilize radio frequencies that do not interfere with or disrupt pacemakers, defibrillators, and any other medical equipment and that receive only signals initiated from the manufacturer ' s system. This REQUIREMENT is not met as evidenced by: K.A.R. 26-40-302 (g) (i) (ii) (iii) The facility reported a census of 95 residents with 5 neighborhoods. Based on observation, interview and record review the facility failed to ensure the wireless call system escalated within in 3 minutes if not answered.	S 970		

LABORATORY DIRECTOR'S OR PROVIDER/SUPPLIER REPRESENTATIVE'S SIGNATURE

TITLE

(X6) DATE

Kansas Department on Aging

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION		(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER: N089001	(X2) MULTIPLE CONSTRUCTION A. BUILDING: _____ B. WING: _____	(X3) DATE SURVEY COMPLETED 02/25/2015
NAME OF PROVIDER OR SUPPLIER BREWSTER HEALTH CENTER		STREET ADDRESS, CITY, STATE, ZIP CODE 1001 SW 29TH ST TOPEKA, KS 66611		
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (EACH DEFICIENCY MUST BE PRECEDED BY FULL REGULATORY OR LSC IDENTIFYING INFORMATION)	ID PREFIX TAG	PROVIDER'S PLAN OF CORRECTION (EACH CORRECTIVE ACTION SHOULD BE CROSS-REFERENCED TO THE APPROPRIATE DEFICIENCY)	(X5) COMPLETE DATE
S 970	<p>Continued From page 1</p> <p>Findings included:</p> <p>- Observation on 2/18/15 during call light checks from 8:15 A.M. - 8:45 A.M., revealed the wireless call system pagers alerted direct care staff when a resident activated their call light. Subsequent recalls alerted the direct care staff again every two minutes before escalating any other staff.</p> <p>An interview on 2/19/15 at 3:55 P.M., Licensed staff I confirmed that the call system alerts the direct care staff immediately and then in two minute increments, after 6 minutes the nurse pager was alerted and the nurse 's station telephone.</p> <p>An interview on 2/23/15 at 4:50 P.M., administrative nurse D revealed the pager system rang in two minutes increments to the direct care staff and after 7 minutes it rang to the charge nurse pager.</p> <p>The facilities Nurse Alert System policy with a revision date of 9/16/14 revealed that on activation of a call light an alert goes to the Kaizen (direct care staff) pagers every two minutes until it is answered and cleared. At 7 minutes the nurse phone gets an alert, and at 9 minutes the Assistant Director of Nursing, managers, Director of Nursing and administrators get alerted by email.</p> <p>The facility failed to ensure the pager system notified additional staff at 3 minute intervals when a call light was activated and not answered.</p>	S 970		